



CERTIFICATE OF ANALYSIS

Sodium Eluant (pH 3.15) with 5% Sulfolane For Amino Acids Analysis

1700-0112

Product: Sodium Eluant pH 3.15 with 5% Sulfolane

Part Number: 1700-0112

Lot Number: 407010

Production Date: 08 Jul 2014

Expiration Date: 08 Jul 2016

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 3.15 ± 0.01 @ 25 ± 0.4 °C (Y/N)

y

2. Resolution meets specifications:

- Threonine/Serine (Y/N)

y

3. Elution time of L-Aspartic acid meets specifications (Y/N)

y

4. Baseline: no peaks (Y/N)

y

Sample: Collagen hydrolysate, **PN:** 012506C

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy
President

Title

Date: 12 Sep 2014