



CERTIFICATE OF ANALYSIS

Artificial Perspiration, AATCC 1700-0012

Product: Artificial Perspiration according to AATCC Test Method 15

Part Number: 1700-0012

Lot Number: 901044

Production Date: 23 Jan 2019

Expiration Date: 23 Jan 2020

Primary Application: Determination of color fastness to perspiration (fabric)

Acceptance Criteria:

1. pH = 4.3 ± 0.1 @ 25 ± 0.4 °C (Y/N)

y

Storage Conditions: Store frozen

Analysis Certified by: Jim Murphy

President

Title

Date:

January 24, 2019

Rev. A, 03/25/14