



Certificate of Analysis

Product Artificial Perspiration, AATCC - Stabilized
Part Number 1700-0015
Primary Application Determination of colorfastness to perspiration
Volume 200 mL
Lot Number 2201046
Production Date 26 January 2022
Expiration Date 26 January 2023
Storage Conditions Dry at 8° to 25°C
Acceptance Criteria

1. Appearance:.....PASS
2. pH = 4.3 ± 0.1 @ $25 \pm 0.4^{\circ}\text{C}$:.....PASS
3. Other components gravimetrically-certified:.....PASS

Components


DI Water	Sodium Phosphate Dibasic
Sodium Chloride	L-Histidine
DL-Lactic Acid	Preservative

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: Maria Ofitserova, Ph.D.

Date: 31 January 2022

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED