



Certificate of Analysis

Product Artificial Urine for Corrosion Testing - Stabilized
Part Number 1700-0016
Primary Application Artificial Urine for corrosion testing of urological implants
Volume 200 mL
Lot Number 2309006
Production Date 05 Sep 2023
Expiration Date 05 Sep 2024
Storage Conditions Dry at 8° to 25°C
Acceptance Criteria

1. Appearance:.....PASS
 2. pH = 6.0 ± 0.1 @ 25 ± 0.4°C :.....PASS
 3. Other components gravimetrically-certified:.....PASS

Components

| | | | |
|--------------------|-------------------|-------------------|--------------|
| DI Water | Sodium Phosphate | Sodium Sulfate | Preservative |
| Sodium Chloride | Calcium Chloride | Sodium Citrate | |
| Potassium Chloride | Ammonium Chloride | Magnesium Sulfate | |

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: David Mazawa

Date: 06 Sep 2023

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED