



Certificate of Analysis

Product Artificial Urine for Corrosion Testing - Stabilized
Part Number 1700-0016
Primary Application Artificial Urine for corrosion testing of urological implants
Volume 200 mL
Lot Number 2507038
Production Date 25 July 2025
Expiration Date 25 July 2026
Storage Conditions Dry at 8° to 25°C
Acceptance Criteria

- 1. Appearance:.....PASS
- 2. pH = 6.0 ± 0.1 @ 25 ± 0.4°C :.....PASS
- 3. Other components gravimetrically-certified:.....PASS

Components

DI Water	Sodium Phosphate	Sodium Sulfate	Preservative
Sodium Chloride	Calcium Chloride	Sodium Citrate	
Potassium Chloride	Ammonium Chloride	Magnesium Sulfate	

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: David Mazawa

Date: 28 July 2025

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED