PICKERING LABORATORIES

Certificate of Analysis

Product	Artificial Urine DIN EN 1616:1999
Part Number	1700-0017
Primary Application	Testing of urethral catheters
Volume	200 mL
Lot Number	2209020
Production Date	12 Sep 2022
Expiration Date	12 Sep 2023
Storage Conditions	Frozen at -20°C

Acceptance Criteria

1. Appearance: ..... PASS
2. pH = 6.6 ± 0.1 @ 25 ± 0.4°C : ..... PASS
3. Other components gravimetrically-certified: ..... PASS

Components

Urea	Creatinine Hydrochloride
Sodium Chloride	
Sodium Phosphate	
Sodium Sulphite	
Ammonium Chloride	

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: David Mazawa **Date:** 12 Sep 2022

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.

GUARANTEED CHEMISTRY CERTIFIED

1280 Space Park Way | Mountain View CA 94043 | USA | 800-654-3330 | 650-694-6700
www.pickeringtestsolutions.com | support@pickeringlabs.com