



Certificate of Analysis

Product Artificial Urine DIN EN 1616:1999
Part Number 1700-0017
Primary Application Testing of urethral catheters
Volume 200 mL
Lot Number 2601015
Production Date 09 Jan 2026
Expiration Date 09 Jan 2027
Storage Conditions Frozen at -20°C

Acceptance Criteria

1. Appearance:.....PASS
2. pH = 6.6 ± 0.1 @ $25 \pm 0.4^\circ\text{C}$:.....PASS
3. Other components gravimetrically-certified:.....PASS

Components

Water	Creatinine Hydrochloride
Urea	Potassium Phosphate
Sodium Chloride	Ammonium Chloride
Sodium Sulfite	
Sodium Phosphate	

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: David Mazawa

Date: 12 Jan 2026

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED