



CERTIFICATE OF ANALYSIS

**Artificial Urine DIN EN 1616:1999
1700-0017**

Product: Artificial Urine prepared according to DIN EN 1616:1999 standard procedure

Part Number: 1700-0017

Lot Number: 503034

Production Date: 23 Mar 2015

Expiration Date: 23 Mar 2016

Primary Application: For testing of urethral catheters.

Acceptance Criteria:

1. pH = 6.6 ± 0.1 @ 25 ± 0.4 °C (Y/N)

y

Storage Conditions: Keep refrigerated

Analysis Certified by: Jim Murphy

President

Title

Date: 24 Mar 2015

Rev. A, 03/26/14