



CERTIFICATE OF ANALYSIS

Artificial Urine DIN EN 1616:1999 1700-0017

Product: Artificial Urine prepared according to DIN EN 1616:1999 standard procedure

Part Number: 1700-0017

Lot Number: 810019

Production Date: 15 Oct 2018

Expiration Date: 15 Oct 2019

Primary Application: For testing of urethral catheters.

Acceptance Criteria:

1. pH = 6.6 ± 0.1 @ 25 ± 0.4 °C (Y/N)

y

Storage Conditions: Keep refrigerated

Analysis Certified by: Jim Murphy
President
Title

Date: October 16, 2018

Rev. A, 03/26/14