



CERTIFICATE OF ANALYSIS

Artificial Urine DIN EN 1616:1999 1700-0017

Product: Artificial Urine prepared according to DIN EN 1616:1999 standard procedure

Part Number: 1700-0017

Lot Number: 911040

Production Date: 25 Nov 2019

Expiration Date: 25 Nov 2020

Primary Application: For testing of urethral catheters.

Acceptance Criteria:

1. pH = 6.6 ± 0.1 @ 25 ± 0.4 °C (Y/N)

y

Storage Conditions: Keep frozen

Analysis Certified by: Jim Murphy
President
Title

Date: November 26,2019

Rev. B, 05/08/19