



## Certificate of Analysis

**Product** Artificial Urine Medium (Non-Sterile)  
**Part Number** 1700-0018  
**Primary Application** Product testing and medical research  
**Volume** 200 mL  
**Lot Number** 2602043  
**Production Date** 19 Feb 2026  
**Expiration Date** 19 Feb 2027  
**Storage Conditions** Frozen at -20°C

### Acceptance Criteria

1. Appearance:.....PASS  
2. pH =  $6.5 \pm 0.1$  @  $25 \pm 0.4^\circ\text{C}$  :.....PASS  
3. Other components gravimetrically-certified:.....PASS

### Components

D.I. Water	Creatinine	Citric Acid	Magnesium Sulfate
Sodium Chloride	Calcium Chloride	Peptone	
Sodium Bicarbonate	Potassium Phosphate	Yeast Extract	
Urea	Ammonium Chloride	Iron Sulfate	
Uric Acid	Lactic Acid	Sodium Sulfate	

\*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

**Analysis Certified By:** David Mazawa

**Date:** 23 Feb 2026

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED