



## CERTIFICATE OF ANALYSIS

### Artificial Eccrine Perspiration with Custom pH – Stabilized 1700-0021

**Product:** Artificial Eccrine Perspiration with custom pH – Stabilized

**Part Number:** 1700-0021

**Lot Number:** 412012

**Production Date:** 09 Dec 2014

**Expiration Date:** 09 Dec 2015

**Primary Application:** Artificial Eccrine Perspiration is a solution of amino acids, salts and metabolites that are found in human eccrine perspiration. It is used in sweat testing of products.

#### Acceptance Criteria:

1. pH = 8.0 ± 0.1 @ 25 ± 0.4 °C (Y/N) y
2. Concentrations of amino acids and other components meet specifications (Y/N): y

**Analysis Certified by:** Jim Murphy **Date:** 19 Dec 2014  
President  
Title