



## CERTIFICATE OF ANALYSIS

### Artificial Eccrine Perspiration – Not Stabilized 1700-0022

**Product:** Artificial Eccrine Perspiration – Not Stabilized

**Part Number:** 1700-0022

**Lot Number:** 003057

**Production Date:** March 30, 2020

**Expiration Date:** March 30, 2021

**Primary Application:** Artificial Eccrine Perspiration is a solution of amino acids, salts and metabolites that are found in human eccrine perspiration. It is used in sweat testing of products.

#### Acceptance Criteria:

1. pH =  $4.5 \pm 0.1$  @  $25 \pm 0.4$  °C (Y/N)

Y

2. Appropriate peaks for amino acids in the test chromatogram (Y/N):

Y

**Analysis Certified by:** Rebecca Smith  
Vice President

Title

**Date:** April 17, 2020