



CERTIFICATE OF ANALYSIS

Artificial Eccrine Perspiration – Not Stabilized 1700-0022

Product: Artificial Eccrine Perspiration – Not Stabilized

Part Number: 1700-0022

Lot Number: 908015

Production Date: 07 Aug 2019

Expiration Date: 07 Aug 2020

Primary Application: Artificial Eccrine Perspiration is a solution of amino acids, salts and metabolites that are found in human eccrine perspiration. It is used in sweat testing of products.

Acceptance Criteria:

1. pH = 4.5 ± 0.1 @ 25 ± 0.4 °C (Y/N) y
2. Appropriate peaks for amino acids in the test chromatogram (Y/N): y

Analysis Certified by: Jim Murphy **Date:** August 15, 2019
President
Title