



## CERTIFICATE OF ANALYSIS

### Artificial Eccrine Perspiration with Custom pH – Not Stabilized 1700-0023

**Product:** Artificial Eccrine Perspiration with custom pH – Not Stabilized

**Part Number:** 1700-0023

**Lot Number:** 211018

**Production Date:** 20 Nov 2012

**Expiration Date:** 20 Nov 2015

**Primary Application:** Artificial Eccrine Perspiration is a solution of amino acids, salts and metabolites that are found in human eccrine perspiration. It is used in sweat testing of products.

#### Acceptance Criteria:

1. pH = 5.0 ± 0.1 @ 25 ± 0.4 °C (Y/N) y
2. Concentrations of amino acids and other components meet specifications (Y/N): y

**Analysis Certified by:** Jim Murphy **Date:** 05 Dec 2012  
President  
Title