



CERTIFICATE OF ANALYSIS

Artificial Eccrine Perspiration with Custom pH – Not Stabilized 1700-0023

Product: Artificial Eccrine Perspiration with custom pH – Not Stabilized

Part Number: 1700-0023

Lot Number: 807001

Production Date: 05 Jul 2018

Expiration Date: 05 Jul 2019

Primary Application: Artificial Eccrine Perspiration is a solution of amino acids, salts and metabolites that are found in human eccrine perspiration. It is used in sweat testing of products.

Acceptance Criteria:

1. pH = 6.0 ± 0.1 @ 25 ± 0.4 °C (Y/N)

y

2. Concentrations of amino acids and other components meet specifications (Y/N):

y

Analysis Certified by: Jim Murphy

Date: July 5, 2018

President

Title