



CERTIFICATE OF ANALYSIS

Artificial Eccrine Perspiration with Custom pH – Not Stabilized 1700-0023

Product: Artificial Eccrine Perspiration with custom pH – Not Stabilized

Part Number: 1700-0023

Lot Number: 906039

Production Date: 21 Jun 2019

Expiration Date: 21 Jun 2020

Primary Application: Artificial Eccrine Perspiration is a solution of amino acids, salts and metabolites that are found in human eccrine perspiration. It is used in sweat testing of products.

Acceptance Criteria:

1. pH = 4.5 ± 0.1 @ 25 ± 0.4 °C (Y/N) y
2. Concentrations of amino acids and other components meet specifications (Y/N): y

Analysis Certified by: Jim Murphy **Date:** June 26, 2019
President
Title