



CERTIFICATE OF ANALYSIS

ChlorAC

1700-0025

Product: ChorAC

Part Number: 1700-0025

Lot Number: 2304028

Production Date: 19 April 2023

Expiration Date: 19 April 2024

Primary Application: Preservative for water samples in Post-column analysis of carbamate pesticides

Acceptance Criteria:

1. pH = 3.0 ± 0.2 @ 25 ± 0.4 °C (Y/N)
(for 3% solution of ChlorAc)
2. No interfering peaks in blank (Y/N)
3. Twelve appropriate peaks in chromatogram (Y/N)

Y
Y
Y

Sample: Carbamate Test Mixture, **PN:** 1700-0063

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: David Mazawa
Title: Technical Support Chemist

Date: 27 April 2023