CERTIFICATE OF ANALYSIS

ChlorAC

1700-0025

Product: ChorAC
Part Number: 1700-0025
Lot Number: 705038
Production Date: 23 May 2017
Expiration Date: 23 May 2018

Primary Application: Preservative for waters samples in Post-column analysis of carbamate pesticides

Acceptance Criteria:

1.	pH = 3.0 ± 0.2 @ 25 ± 0.4 °C (Y/N) (for 3% solution of ChlorAc)	**y**
2.	No interfering peaks in blank (Y/N)	**y**
3.	Twelve appropriate peaks in chromatogram (Y/N)	**y**

Sample: Carbamate Test Mixture, **PN:** 1700-0063

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: Jim Murphy **Date:** June 5, 2017
President
Title

1280 Space Park Way · Mountain View CA 94043 · USA
800-654-3330 · 650-694-6700 · 650-968-0749 FAX · www.pickeringlabs.com · support@pickeringlabs.com