



CERTIFICATE OF ANALYSIS

Artificial Perspiration, ISO 3160

1700-0026

Product: Artificial Perspiration according to ISO 3160

Part Number: 1700-0026

Lot Number: 010022

Production Date: October 13, 2020

Expiration Date: October 13, 2021

Primary Application: Determination of corrosion resistance

Acceptance Criteria:

1. pH = 4.7 ± 0.04 @ 25 ± 0.4 °C (Y/N)

Y

Storage Conditions: Store frozen

Analysis Certified by: Jim Murphy
President

Title

Date: October 27, 2020