



## Certificate of Analysis

**Product** Artificial Perspiration, ISO 3160, Not Stabilized  
**Part Number** 1700-0026  
**Primary Application** Determination of corrosion resistance  
**Volume** 200 mL  
**Lot Number** 2503047  
**Production Date** 27 Mar 2025  
**Expiration Date** 27 Mar 2026  
**Storage Conditions** Frozen at -20°C

### Acceptance Criteria

- 1. Appearance:.....PASS
- 2. pH = 4.7±0.04 @ 25±0.4°C :.....PASS
- 3. Other components gravimetrically-certified:.....PASS

### Components

DI Water	Urea	Sodium Hydroxide
Sodium Chloride	Acetic Acid	
Ammonium Chloride	Lactic Acid	

\*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

**Analysis Certified By:** David Mazawa

**Date:** 31 Mar 2025

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED