



Certificate of Analysis

Product Simulated Body Fluid, ISO 23317-14
Part Number 1700-0028
Primary Application Product testing and medical research
Volume 200 mL
Lot Number 2601014
Production Date 19 Jan 2026
Expiration Date 19 Jan 2027
Storage Conditions Refrigerated at 4-8°C

Acceptance Criteria

1. Appearance:.....PASS
2. pH = 7.40 ± 0.01 @ 36.5 ± 0.1 °C :.....PASS
3. Other components gravimetrically-certified:.....PASS

Components

D.I. Water	Hydrochloric Acid
Sodium Chloride	Calcium Chloride
Sodium Bicarbonate	Potassium Phosphate
Magnesium Chloride	Sodium Sulfate
Potassium Chloride	Tris

*All product components are weighed on NIST-traceable scales and measured in Class A volumetric glassware

Analysis Certified By: David Mazawa

Date: 20 Jan 2026

The product is manufactured, packaged, stored and shipped according to manufacturing standard best practices and under modern sanitary conditions.



CERTIFIED