



CERTIFICATE OF ANALYSIS

Carbamate Test Mixture

1700-0063

Product: Carbamate Test Mixture

Part Number: 1700-0063

Lot Number: 711021

Production Date: 09 Nov 2017

Expiration Date: 09 Nov 2020

Primary Application: Carbamate Test Mixture is a solution of 12 compounds in methanol. Concentration of each compound is 2.5 µg/mL. This Test Mixture is intended to be used for establishing an elution profile and for troubleshooting. This product is not intended for calibration.

Acceptance Criteria:

1. The concentrations of Carbamate pesticides and Naphthol are within +/- 10% (Y/N)

y

Storage Conditions: Keep frozen

Analysis Certified by: Jim Murphy **Date:** Nov. 25, 2017
President
Title