



## CERTIFICATE OF ANALYSIS

### Carbamate Test Mixture

1700-0063

**Product:** Carbamate Test Mixture

**Part Number:** 1700-0063

**Lot Number:** 909007

**Production Date:** 23 Sept 2019

**Expiration Date:** 22 Sept 2022

**Primary Application:** Carbamate Test Mixture is a solution of 12 compounds in methanol. Concentration of each compound is 2.5 µg/mL. This Test Mixture is intended to be used for establishing an elution profile and for troubleshooting. This product is not intended for calibration.

#### Acceptance Criteria:

1. The concentrations of Carbamate pesticides and Naphthol are within +/- 10% (Y/N)

y

**Storage Conditions:** Keep frozen

**Analysis Certified by:** Jim Murphy **Date:** October 22, 2019  
President  
Title