



CERTIFICATE OF ANALYSIS

Sodium Eluant (pH 3.15) with 5% Sulfolane for amino Acids Analysis

1700-0112

Product: Sodium Eluant pH 3.15 with 5% Sulfolane

Part Number: 1700-0112

Lot Number: 109007

Production Date: 09 Sep 2011

Expiration Date: 07 Sep 2013

Primary Application: Sodium based amino acids analysis

Acceptance Criteria

1. pH = 3.15 ± 0.01 @ 25 ± 0.4 °C (Y/N)
2. Retention time of L-aspartic acid meets specifications. (Y/N)
3. Resolution of threonine/serine $\geq 93\%$ (Y/N)
4. No peaks in baseline (Y/N)

Y
Y
Y
Y

Sample: Collagen Hydrolysate, PN: 012506C,

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: _____

[Signature]
VP OPR

Title

Date: 20 Sep 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 8, 03/28/2011