



## CERTIFICATE OF ANALYSIS

**Sodium Eluant (pH 3.15) with 5% Sulfolane for amino Acids Analysis**

**1700-0112**

**Product:** Sodium Eluant pH 3.15 with 5% Sulfolane

**Part Number:** 1700-0112

**Lot Number:** 110005

**Production Date:** 07 Oct 2011

**Expiration Date:** 06 Oct 2013

**Primary Application:** Sodium based amino acids analysis

### Acceptance Criteria

1. pH =  $3.15 \pm 0.01$  @  $25 \pm 0.4$  °C (Y/N)
2. Retention time of L-aspartic acid meets specifications. (Y/N)
3. Resolution of threonine/serine  $\geq 93\%$  (Y/N)
4. No peaks in baseline (Y/N)

Y  
Y  
Y  
Y

**Sample:** Collagen Hydrolysate, PN: 012506C,

**Storage Conditions:** Dry at 8°C to 25°C

**Analysis Certified by:** \_\_\_\_\_

*[Signature]*  
VP OPR  
Title

**Date:** 19 Oct 2011

# ACCEPTED

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 8, 03/28/2011