



CERTIFICATE OF ANALYSIS

**Sodium Eluant (pH 3.15) with 5% Sulfolane
For Amino Acids Analysis**

1700-0112

Product: Sodium Eluant pH 3.15 with 5% Sulfolane

Part Number: 1700-0112

Lot Number: 203004

Production Date: 06 MAR 2012

Expiration Date: 05 MAR 2014

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 3.15 ± 0.01 @ 25 ± 0.4 °C (Y/N)
2. Resolution meets specifications:
 - Threonine/Serine (Y/N)
3. Elution time of L-Aspartic acid meets specifications (Y/N)
4. Baseline: no peaks (Y/N)

Y
Y
Y
Y

Sample: Collagen hydrolysate, PN: 012506C

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: 

Date: 19 MAR 2012

Title

ACCEPTED

Rev. 09, 11/09/2011

PICKERING LABORATORIES, INC. P:\Documentation\Chemical Documentation\1700-0112
QUALITY ASSURANCE