



CERTIFICATE OF ANALYSIS

Sodium Eluant (pH 3.15) with 5% Sulfolane For Amino Acids Analysis

1700-0112

Product: Sodium Eluant pH 3.15 with 5% Sulfolane

Part Number: 1700-0112

Lot Number: 2510032

Production Date: 14 Oct 2025

Expiration Date: 14 Oct 2027

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 3.15 ± 0.01 @ 25 ± 0.4 °C (Y/N)	<u>Y</u>
2. Resolution meets specifications:	
• Threonine/Serine (Y/N)	<u>Y</u>
3. Elution time of L-Aspartic acid meets specifications (Y/N)	<u>Y</u>
4. Baseline: no peaks (Y/N)	<u>Y</u>

Sample: Collagen hydrolysate, **PN:** 012506C

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: David Mazawa **Date:** 20 Oct 2025
Title: QC Manager