PICKERING LABORATORIES

# CERTIFICATE OF ANALYSIS

## Sodium Eluant (pH 3.15) with 5% Sulfolane For Amino Acids Analysis

## 1700-0112

**Product:** Sodium Eluant pH 3.15 with 5% Sulfolane
**Part Number:** 1700-0112
**Lot Number:** 606018
**Production Date:** 14 Jun 2016
**Expiration Date:** 14 Jun 2018

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

| | |
|---|---|
| 1. pH = 3.15 ± 0.01 @ 25 ± 0.4 °C (Y/N) | **y** |
| 2. Resolution meets specifications: | |
| • Threonine/Serine (Y/N) | **y** |
| 3. Elution time of L-Aspartic acid meets specifications (Y/N) | **y** |
| 4. Baseline: no peaks (Y/N) | **y** |

**Sample:** Collagen hydrolysate, **PN**: 012506C

**Storage Conditions: Dry at 8 to 25 °C**

**Analysis Certified by:** Jim Murphy **Date:** June 27, 2016
President
**Title**



1280 Space Park Way · Mountain View CA 94043 · USA
800-654-3330 · 650-694-6700 · 650-968-0749 FAX · www.pickeringlabs.com · support@pickeringlabs.com