



CERTIFICATE OF ANALYSIS

**Sodium Eluant pH 3.33
With 2.0 % Sulfolane
1700-0114**

Product: Sodium Eluant pH 3.33 with 2.0 % Sulfolane

Part Number: 1700-0114

Lot Number: 103017

Production Date: 16 MAR 2011

Expiration Date: 15 Sep 2012

Primary Application: Sodium-based amino acid analysis

Acceptance Criteria:

1. pH = 3.28 ± 0.01 @ 25 ± 0.4 °C (Y/N)
2. Peak resolution meets specifications(Y/N)
3. No peaks in baseline (Y/N)

Y
Y
Y

Calibration Standard: Protien Hydrolysate: PN 012506H,

Analysis Certified by: _____

Jim V. [Signature]
VP OPR

Date: 21 MAR 2011

Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 9, 03/28/2011