



## CERTIFICATE OF ANALYSIS

### Sodium Eluant With 2.0 % Sulfolane 1700-0114

**Product:** Sodium Eluant with 2.0 % Sulfolane

**Part Number:** 1700-0114

**Lot Number:** 212010

**Production Date:** 07 Dec 2012

**Expiration Date:** 07 Dec 2014

**Primary Application:** Sodium-based amino acid analysis

#### Acceptance Criteria:

1. pH =  $3.28 \pm 0.01$  @  $25 \pm 0.4$  °C (Y/N)

y

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy

President

Title

**Date:** 21 Dec 2012