



## CERTIFICATE OF ANALYSIS

**ChlorAC**

**1700-0132**

**Product:** ChorAC  
**Part Number:** 1700-0132  
**Lot Number:** 006028  
**Production Date:** June 12, 2020  
**Expiration Date:** June 12, 2021

**Primary Application:** Preservative for waters samples in Post-column analysis of carbamate pesticides

**Acceptance Criteria:**

- |   |          |
|---|----------|
| 1. pH = $3.0 \pm 0.2$ @ $25 \pm 0.4$ °C (Y/N)<br>(for 3% solution of ChlorAc) | <u>Y</u> |
| 2. No interfering peaks in blank (Y/N)  | <u>Y</u> |
| 3. Twelve appropriate peaks in chromatogram (Y/N)                             | <u>Y</u> |

**Sample:** Carbamate Test Mixture, **PN:** 1700-0063

**Storage Conditions:** Dry at 8°C to 25°C

**Analysis Certified by:** Rebecca Smith **Date:** August 13, 2020  
Vice President  
Title