



## CERTIFICATE OF ANALYSIS

ChlorAC

1700-0132

Product: ChorAC

Part Number: 1700-0132

Lot Number: 1707035

Test Start Date: 29 Jul 2010

Primary Application: Post-column analysis of carbamate pesticides

**Acceptance Criteria:**

1. pH =  $3.0 \pm 0.2$  @  $25 \text{ }^\circ\text{C} \pm 0.4 \text{ }^\circ\text{C}$  (for a 3% solution of ChlorAC).
2. No interfering peaks in blank.
3. There are 12 peaks with qualitative Carbamate Test Standard (1700-0063).

Calibration Standard: Carbamate Test Standard, PN 1700-0063, LN 908045

Analysis Certified by: *Jim Murphy* Date: 04 Aug 2010

VP Ops

Title

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE