



CERTIFICATE OF ANALYSIS

ChlorAC

1700-0132

Product: ChorAC

Part Number: 1700-0132

Lot Number: 012026

Test Start Date: 16 DEC 2010

Primary Application: Post-column analysis of carbamate pesticides

Acceptance Criteria:

1. pH = 3.0 ± 0.2 @ $25 \text{ }^\circ\text{C} \pm 0.4 \text{ }^\circ\text{C}$ (for a 3% solution of ChlorAC).
2. No interfering peaks in blank.
3. There are 12 peaks with qualitative Carbamate Test Standard (1700-0063).

Calibration Standard: Carbamate Test Standard, PN 1700-0063, LN 908045

Analysis Certified by: *Jim Murphy* **Date:** 18 JAN 2011

VP OPR
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE