



CERTIFICATE OF ANALYSIS

ChlorAC

1700-0132

Product: ChorAC

Part Number: 1700-0132

Lot Number: 104005

Production Date: 08 APR 2011

Expiration Date: 08 APR 2012

Primary Application: Preservative for waters samples in Post-column analysis of carbamate pesticides

Acceptance Criteria:

- 1. pH = 3.0 ± 0.2 @ 25 ° C ± 0.4 ° C (Y/N) y
- (for a 3% solution of ChlorAC)
- 2. No interfering peaks in blank (Y/N) y
- 3. 12 appropriate peaks in the chromatogram (Y/N) y

Sample: Carbamate Test Mixture, PN: 1700-0063

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: *Jim Murphy* Date: 18 Apr 2011
VP Ops
 Title

ACCEPTED
 PICKERING LABORATORIES, INC.
 QUALITY ASSURANCE