



## CERTIFICATE OF ANALYSIS

**ChlorAC**

**1700-0132**

**Product:** ChorAC  
**Part Number:** 1700-0132  
**Lot Number:** 104037  
**Production Date:** 04 MAY 2011  
**Expiration Date:** 02 MAY 2012

**Primary Application:** Preservative for waters samples in Post-column analysis of carbamate pesticides

**Acceptance Criteria:**

1. pH =  $3.0 \pm 0.2$  @  $25 \text{ }^\circ\text{C} \pm 0.4 \text{ }^\circ\text{C}$  (Y/N) 4
2. No interfering peaks in blank (Y/N) 4
3. 12 appropriate peaks in the chromatogram (Y/N) 4

**Sample:** Carbamate Test Mixture, **PN:** 1700-0063

**Storage Conditions:** Dry at  $8^\circ\text{C}$  to  $25^\circ\text{C}$

**Analysis Certified by:** *Jim Murphy* **Date:** 05 MAY 2011  
*V.P. Ope*  
**Title**

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 5, 04/08/2011