



CERTIFICATE OF ANALYSIS

ChlorAC

1700-0132

Product: ChorAC
Part Number: 1700-0132
Lot Number: 106016
Production Date: 20 JUN 2011
Expiration Date: 17 JUN 2012


Primary Application: Preservative for waters samples in Post-column analysis of carbamate pesticides

Acceptance Criteria:

1. pH = 3.0 ± 0.2 @ $25^\circ\text{C} \pm 0.4^\circ\text{C}$ (Y/N) y
2. No interfering peaks in blank (Y/N) y
3. Twelve appropriate peaks in the chromatogram (Y/N) y

Sample: Carbamate Test Mixture, **PN:** 1700-0063

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by:  Date: 21 JUN 2011
VP OPR
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE