



CERTIFICATE OF ANALYSIS

ChlorAC

1700-0132

Product: ChorAC

Part Number: 1700-0132

Lot Number:

202005

Production Date:

02 Feb 2012

Expiration Date:

02 Feb 2013

Primary Application: Preservative for waters samples in Post-column analysis of carbamate pesticides

Acceptance Criteria:

1. pH = 3.0 ± 0.2 @ 25 ± 0.4 °C (Y/N)
(for 3% solution of ChlorAc)
2. No interfering peaks in blank (Y/N)
3. Twelve appropriate peaks in chromatogram (Y/N)

4
4
4

Sample: Carbamate Test Mixture, PN: 1700-0063

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by:

[Signature]
Title

Date:

19 MAR 2012

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

P:\Documentation\Chemical Documentation\CAc 1700-0132

Rev. A, 06/17/2011