



CERTIFICATE OF ANALYSIS

ChlorAC

1700-0132

Product: ChorAC

Part Number: 1700-0132

Lot Number: 212016

Production Date: 14 Dec 2012

Expiration Date: 14 Dec 2013

Primary Application: Preservative for waters samples in Post-column analysis of carbamate pesticides

Acceptance Criteria:

- | | |
|---|----------|
| 1. pH = 3.0 ± 0.2 @ 25 ± 0.4 °C (Y/N)
(for 3% solution of ChlorAc) | <u>y</u> |
| 2. No interfering peaks in blank (Y/N) | <u>y</u> |
| 3. Twelve appropriate peaks in chromatogram (Y/N) | <u>y</u> |

Sample: Carbamate Test Mixture, **PN:** 1700-0063

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: Jim Murphy **Date:** 212 Dec 2012
President
Title