



CERTIFICATE OF ANALYSIS

ChlorAC

1700-0132

Product: ChorAC

Part Number: 1700-0132

Lot Number: 2203024

Production Date: March 11, 2022

Expiration Date: March 11, 2023

Primary Application: Preservative for water samples in Post-column analysis of carbamate pesticides

Acceptance Criteria:

- | | |
|---|----------|
| 1. pH = 3.0 ± 0.2 @ 25 ± 0.4 °C (Y/N)
(for 3% solution of ChlorAc) | <u>Y</u> |
| 2. No interfering peaks in blank (Y/N) | <u>Y</u> |
| 3. Twelve appropriate peaks in chromatogram (Y/N) | <u>Y</u> |

Sample: Carbamate Test Mixture, **PN:** 1700-0063

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: Jim Murphy **Date:** March 22, 2022
President
Title