



CERTIFICATE OF ANALYSIS

ChlorAC

1700-0132

Product: ChorAC
Part Number: 1700-0132
Lot Number: 404041
Production Date: 01 May 2014
Expiration Date: 01 May 2015

Primary Application: Preservative for waters samples in Post-column analysis of carbamate pesticides

Acceptance Criteria:

1. pH = 3.0 ± 0.2 @ 25 ± 0.4 °C (Y/N) (for 3% solution of ChlorAc) y
2. No interfering peaks in blank (Y/N) y
3. Twelve appropriate peaks in chromatogram (Y/N) y

Sample: Carbamate Test Mixture, **PN:** 1700-0063

Storage Conditions: Dry at 8°C to 25°C

Analysis Certified by: Jim Murphy **Date:** 08 May 2014
President
Title