



# CERTIFICATE OF ANALYSIS

RESTORE

1700-0140

Product: RESTORE  
Part Number: 1700-0140  
Lot Number: 007004  
Test Start Date: 07 Jul 2010


**Primary Application:** Removes metal ion contamination from ion-exchange guard or analytical columns in glyphosate analysis.

**Acceptance Criteria:**

1. pH =  $1.30 \pm 0.05$  @  $25 \pm 0.4$  °C

**Actual Values:**

- 1.318

Analysis Certified by:   
VP Opp  
Title

Date: 15 Jul 2010

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 4, 1/29/04

1280 Space Park Way · Mountain View CA 94043 · USA  
800-654-3330 · 650-694-6700 · 650-968-0749 FAX · [www.pickeringlabs.com](http://www.pickeringlabs.com) · [support@pickeringlabs.com](mailto:support@pickeringlabs.com)