



# CERTIFICATE OF ANALYSIS

**RESTORE**

**1700-0140**

**Product:** RESTORE  
**Part Number:** 1700-0140  
**Lot Number:** 01203G  
**Test Start Date:** 20 DEC 2010

**Primary Application:** Removes metal ion contamination from ion-exchange guard or analytical columns in glyphosate analysis.

**Acceptance Criteria:**

**Actual Values:**

1. pH =  $1.30 \pm 0.05$  @  $25 \pm 0.4$  °C

1.294

**Analysis Certified by:** \_\_\_\_\_

*Jim Murphy*  
VP OPR

**Date:** \_\_\_\_\_

18 JAN 2011

Title

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE