



CERTIFICATE OF ANALYSIS

RESTORE

1700-0140

Product: RESTORE

Part Number: 1700-0140

Lot Number: 101031

Test Start Date: 24 JAN 2011

Primary Application: Removes metal ion contamination from ion-exchange guard or analytical columns in glyphosate analysis.

Acceptance Criteria:

Actual Values:

1. pH = 1.30 ± 0.05 @ 25 ± 0.4 °C

- 1.258

Analysis Certified by: _____

Jim Myhr

Date: 15 Feb 2011

V.P. OPR

Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE