



# CERTIFICATE OF ANALYSIS

**RESTORE**

**1700-0140**

**Product:** RESTORE  
**Part Number:** 1700-0140  
**Lot Number:** 110017  
**Production Date:** 13 Oct 2011  
**Expiration Date:** 13 Oct 2012

**Primary Application:** Removes metal ion contamination from ion-exchange guard or analytical columns in glyphosate analysis.

**Acceptance Criteria:**

1. pH =  $1.30 \pm 0.05$  @  $25 \pm 0.4$  °C (Y/N)

4

Analysis Certified by:

*Jim Murphy*  
VP OPR  
Title

Date: 24 Oct 2011

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE