



CERTIFICATE OF ANALYSIS

RESTORE

1700-0140

Product: RESTORE

Part Number: 1700-0140

Lot Number: 209032

Production Date: 24 Sep 2012

Expiration Date: 24 Sep 2014

Primary Application: Removes metal ion contamination from ion-exchange guard or analytical columns in glyphosate analysis.

Acceptance Criteria:

1. pH = 1.30 ± 0.05 @ 25 ± 0.4 °C (Y/N)

y

Analysis Certified by: Jim Murphy

Date: 05 Oct 2012

VP Operations

Title