



## CERTIFICATE OF ANALYSIS

**RESTORE**

**1700-0140**

**Product:** RESTORE

**Part Number:** 1700-0140

**Lot Number:** 2305041

**Production Date:** 26 May 2023

**Expiration Date:** 26 May 2025

**Primary Application:** Removes metal ion contamination from ion-exchange guard or analytical columns in glyphosate analysis.

**Acceptance Criteria:**

1. pH =  $1.30 \pm 0.05$  @  $25 \pm 0.4$  °C (Y/N)

Y

**Analysis Certified by:** David Mazawa

**Date:** 02 June 2023

**Title:** Technical Support Chemist