



# CERTIFICATE OF ANALYSIS

**RESTORE**

**1700-0140**

**Product:** RESTORE

**Part Number:** 1700-0140

**Lot Number:** 2410035

**Production Date:** 17 Oct 2024

**Expiration Date:** 17 Oct 2026

**Primary Application:** Removes metal ion contamination from ion-exchange guard or analytical columns in glyphosate analysis.

**Acceptance Criteria:**

1. pH =  $1.30 \pm 0.05$  @  $25 \pm 0.4$  °C (Y/N) Y

**Analysis Certified by:** David Mazawa **Date:** 22 Oct 2024

**Title:** QC Manager