



CERTIFICATE OF ANALYSIS

RESTORE

1700-0140

Product: RESTORE

Part Number: 1700-0140

Lot Number: 2506004

Production Date: 02 Jun 2025

Expiration Date: 02 Jun 2027

Primary Application: Removes metal ion contamination from ion-exchange guard or analytical columns in glyphosate analysis.

Acceptance Criteria:

1. pH = 1.30 ± 0.05 @ 25 ± 0.4 °C (Y/N)

Y

Analysis Certified by: David Mazawa
Title: QC Manager

Date: 05 Jun 2025