



CERTIFICATE OF ANALYSIS

RESTORE

1700-0140

Product: RESTORE

Part Number: 1700-0140

Lot Number: 312019

Production Date: 19 Dec 2013

Expiration Date: 19 Dec 2015

Primary Application: Removes metal ion contamination from ion-exchange guard or analytical columns in glyphosate analysis.

Acceptance Criteria:

1. pH = 1.30 ± 0.05 @ 25 ± 0.4 °C (Y/N)

Y

Analysis Certified by: Jim Murphy

Date: 24 Jan 2014

President

Title